Headspace GC-FID Maintenance Log

Instrument: BAC-1 **Month/Year:** January 2025

Date	Nitrogen Tank Pressure	Hydrogen Tank Pressure	Helium Tank Pressure	Air Control Pass/Fail	PC Pass/Fail	Comments	Analyst Initials
1/23/2025	N/A	N/A	N/A	N/A	N/A	PM performed. Both columns replaced.	mh
1/24/2025	1750	900	2000	Pass 1130/2025 mh	Pass	verification He tank changed 1/20/2025.	mh

KB 2/7/2025

Signature: Ralph M. Baylor **Date:** 2/7/2025