Brazoria County Crime Lab

Toxicology Section

Headspace GC-FID Maintenance Log

Instrument: BAC-1 Month/Year: May 2025

Date	Nitrogen Tank Pressure	Hydrogen Tank Pressure	Helium Tank Pressure	Air Control Pass/Fail	PC Pass/Fail	Comments	Analyst Initials
5/7/2025	1150	400	1050	Pass	Pass		KB
5/9/2025	900	400	800	Pass	Pass		KB
5/15/2025	600	400	2400	Pass	Pass	He tank changed 5/12/2025	KB
5/30/2025	2600	400	200	Pass	Pass	Nirtogen tank changed 5/29/2025	KB

Signature: Kayla M. Baylor Digitally signed by Kayla M. Baylor
Date: 2025.06.19 10:25:10 -05'00' Date: 6/15/2025