Brazoria County Crime Lab

Toxicology Section

Headspace GC-FID Maintenance Log

Instrument: BAC-1 **Month/Year:** October 2025

Date	Nitrogen Tank Pressure	Hydrogen Tank Pressure	Helium Tank Pressure	Air Control Pass/Fail	PC Pass/Fail	Comments	Analyst Initials
10/1/2025	750	2000	2100	Pass	Pass		KB
10/2/2025	650	2000	2000	Pass	Pass		KB
10/9/2025	1600	2000	1000	Pass	Pass		KB

Signature: Kayla M. Baylor Digitally signed by Kayla M. Baylor Date: 2025.12.17 07:24:49 -06'00' **Date:** 12/17/2025