Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 5/1/2025 Temp: 20.8 °C
Analyst: DN Humidity: 71 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.01 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	1.00 g	2	5.00 g	5.01 g	2	30.00 g	30.01 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 5/5/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 66 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.01 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.97 g

Date: 5/12/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 63 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.01 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.97 g
Balance Readout 2: 1.01 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Date: 5/19/2025 Temperature: 21.5 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Date: 5/27/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.95 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AP210S

Monthly Check

Date: 5/1/2025 Temp: 20.8 °C
Analyst: DN Humidity: 71 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.0000 g	1 3.00 g	1 3.0000 g	1 5.00 g	1 5.0000 g
2 1.00 g	2 1.0000 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 5.0000 g
3 1.00 g	3 0.9999 g	3 3.00 g	3 2.9999 g	3 5.00 g	3 4.9999 g

Surrogate Checks

Date: 5/5/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 66 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0053 g	1 99.9594 g
2 1.0010 g	2 5.0052 g	2 99.9594 g
3 1.0010 g	3 5.0052 g	3 99.9594 g

Date: 5/12/2025 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: DN Humidity: 63 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0008 g	1 5.0051 g	1 99.9594 g
2 1.0009 g	2 5.0051 g	2 99.9594 g
3 1.0009 g	3 5.0051 g	3 99.9594 g

Date: 5/19/2025 Temperature: 21.5 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9596 g
2 1.0009 g	2 5.0052 g	2 99.9595 g
3 1.0009 g	3 5.0052 g	3 99.9596 g

Date: 5/27/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9593 g
2 1.0010 g	2 5.0052 g	2 99.9593 g
3 1.0009 g	3 5.0051 g	3 99.9592 g