Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 3/2/2026 Temp: 21.3 °C
Analyst: KK Humidity: 69 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	1.01 g	2	5.00 g	5.00 g	2	30.00 g	30.00 g
3	1.00 g	1.01 g	3	5.00 g	5.00 g	3	30.00 g	29.99 g

Surrogate Checks

Date: 3/9/2026 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: KK Humidity: 71 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.02 g	Balance Readout	1 99.96 g
	2 1.01 g		2 5.01 g		2 99.96 g
	3 1.01 g		3 5.01 g		3 99.96 g

Date: 3/16/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 33 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.00 g	Balance Readout	1 99.95 g
	2 1.00 g		2 5.01 g		2 99.96 g
	3 1.00 g		3 5.01 g		3 99.96 g

Date: 3/23/2026 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: KK Humidity: 64 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.01 g	Balance Readout	1 5.01 g	Balance Readout	1 99.97 g
	2 1.01 g		2 5.01 g		2 99.97 g
	3 1.00 g		3 5.01 g		3 99.97 g

Date: 3/30/2026 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 64 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.00 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.00 g		2 99.97 g
	3 1.01 g		3 5.00 g		3 99.97 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification AP210S

-DN 4/2/2026

Monthly Check

Date _____ Temp _____ °C
Analyst _____ Humidity _____ %
Weight Set Used _____

Balance not in service -KK 3/2/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1 _____ g	1	3.00 g	1 _____ g	1	5.00 g	1 _____ g
2	1.00 g	2 _____ g	2	3.00 g	2 _____ g	2	5.00 g	2 _____ g
3	1.00 g	3 _____ g	3	3.00 g	3 _____ g	3	5.00 g	3 _____ g

Surrogate Checks

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g