Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: MPC2000

Monthly Check

Date: 5/1/2026 Temp: 21 °C
Analyst: KK Humidity: 63 %
Weight Set Used: 9072

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.01 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.01 g	2 5.00 g	2 5.01 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.01 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: KK Humidity: 58 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.01 g	3 5.00 g	3 99.97 g

Date: 5/11/2026 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: KK Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.01 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.01 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.97 g

Date: 5/26/2026 Temperature: 21.0 °C Weight Set Used: 155617
Analyst: KK Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification AP210S

Monthly Check

Date _____ Temp _____ °C
Analyst _____ Humidity _____ %
Weight Set Used _____

Notes: balance not in service. -KK 4/14/2026

N/A -KK 6/1/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1 _____ g	1	3.00 g	1 _____ g	1	5.00 g	1 _____ g
2	1.00 g	2 _____ g	2	3.00 g	2 _____ g	2	5.00 g	2 _____ g
3	1.00 g	3 _____ g	3	3.00 g	3 _____ g	3	5.00 g	3 _____ g

N/A -KK 6/1/2026

Surrogate Checks

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g

Date _____ Temperature _____ °C Weight Set Used _____
Analyst _____ Humidity _____ %

Low Weight Used 1.00 g	Mid Weight Used 5.00 g	High Weight Used 100.00 g
Balance Readout 1 _____ g	Balance Readout 1 _____ g	Balance Readout 1 _____ g
2 _____ g	2 _____ g	2 _____ g
3 _____ g	3 _____ g	3 _____ g