Pipette Monitoring and Maintenance

Manufacturer Transferpette Model _____
Serial Number 22H33903 Capacity 2-20µl

Pipette Identification 3903

Date 10/14/2025 Analyst mh H_2O °C Start 20.3 H_2O °C End 20.3 Room °C 20.3 Humidity % 64 Theoretical ρ N/A

Low 10µl		Mid-Level 20µl		High		☑ Pass ☐ Re-Check ☐ Fail
Measured	Actual	Measured	Actual	Measured	Actual	Notes:
1. 10	1. 0.0103	1. 20	1. 0.0199	1.	1.	
2. 10	2. 0.0101	2. 20	2. 0.0206	2. 10/14/2025 mh	2.	
3. 10	3. 0.0105	3. 20	3. 0.0202	3.	3.	
Actual: $\bar{x}$ = 10.3 σ = 0.200 % 103		Actual: $\bar{x}$ = 20.2 σ = 0.351 % 101.2		Actual: $\bar{x}$ = _____ σ = _____ % _____		

Date _____ Analyst _____ H_2O °C Start _____ H_2O °C End _____ Room °C _____ Humidity % _____ Theoretical ρ _____

Low		Mid-Level		High		☐ Pass ☐ Re-Check ☐ Fail
Measured	Actual	Measured	Actual	Measured	Actual	Notes:
1.	1.	1.	1. 10/14/2025 mh	1.	1.	
2.	2.	2.	2.	2.	2.	
3.	3.	3.	3.	3.	3.	
Actual: $\bar{x}$ = _____ σ = _____ % _____		Actual: $\bar{x}$ = _____ σ = _____ % _____		Actual: $\bar{x}$ = _____ σ = _____ % _____		