Brazoria County Crime Laboratory

Reagent Preparation Log

Reagent Name	0.01% n-propanol Interal Standard (% v/v)			**Inventory #**	R-26-5002		
Pipette(s)	N/A		**Balance (SN)**	Analytical:	Top Loading: (N/A)		
Prepared By	KB	**Preparation Date**	1/12/2026		**Expiration Date**	7/12/2026	
Total Volume	2000 mL	**Verification Batch Name**		not used in casework - KB 1/21/2026			
Chemical 1	1% n-propanol Internal Standard Stock Solution	**Manufacturer and Lot #**	In-house R-26-5001	**Expiration Date**	7/12/2026	**Amount**	20mL
Chemical 2	LC Water (C-25-0015I)	**Manufacturer and Lot #**	Thermo Scientific 235286	**Expiration Date**	RT: 6/30/2030	**Amount**	1980mL
Chemical 3		**Manufacturer and Lot #**		**Expiration Date**		**Amount**	
Chemical 4		**Manufacturer and Lot #**	KB 1/21/2026	**Expiration Date**		**Amount**	
Chemical 5		**Manufacturer and Lot #**		**Expiration Date**		**Amount**	
Comments:							