Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Monthly Check

Date: 4/1/2025 Temp: 20.6 °C
Analyst: DN Humidity: 65 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.9999 g	1 3.00 g	1 3.0001 g	1 5.00 g	1 5.0001 g
2 1.00 g	2 1.0000 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 5.0000 g
3 1.00 g	3 1.0001 g	3 3.00 g	3 3.0000 g	3 5.00 g	3 5.0000 g

Surrogate Checks

Date: 4/7/2025 Temperature: 19.8 °C Weight Set Used: 9072
Analyst: DS Humidity: 42 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0000 g	1 4.9999 g	1 99.9999 g
2 0.9999 g	2 4.9999 g	2 100.0001 g
3 1.0002 g	3 5.0001 g	3 100.0001 g

Date: 4/14/2025 Temperature: 20.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9591 g
2 1.0010 g	2 5.0053 g	2 99.9591 g
3 1.0010 g	3 5.0052 g	3 99.9591 g

Date: 4/22/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0054 g	1 99.9597 g
2 1.0011 g	2 5.0052 g	2 99.9595 g
3 1.0010 g	3 5.0053 g	3 99.9595 g

Date: 4/28/2025 Temperature: 21.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0053 g	1 99.9595 g
2 1.0010 g	2 5.0053 g	2 99.9595 g
3 1.0010 g	3 5.0052 g	3 99.9595 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD1

Monthly Check

Date: 4/1/2025 Temp: 20.6 °C
Analyst: DN Humidity: 65 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	1.00 g	2	5.00 g	5.00 g	2	30.00 g	30.00 g
3	1.00 g	1.00 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 4/7/2025 Temperature: 20.0 °C Weight Set Used: 9072
Analyst: DS Humidity: 40 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 0.99 g	Balance Readout 1: 4.99 g	Balance Readout 1: 100.00 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 100.00 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 100.01 g

Date: 4/14/2025 Temperature: 20.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.96 g

Date: 4/22/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.96 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.96 g

Date: 4/28/2025 Temperature: 21.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.95 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD2

Monthly Check

Date: 4/1/2025 Temp: 20.6 °C
Analyst: DN Humidity: 65 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 0.99 g	2 5.00 g	2 4.99 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.00 g	3 5.00 g	3 4.99 g	3 30.00 g	3 29.99 g

Surrogate Checks

Date: 4/7/2025 Temperature: 20.2 °C Weight Set Used: 9072
Analyst: DS Humidity: 39 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.01 g	1 99.99 g
2 1.01 g	2 4.99 g	2 100.00 g
3 1.00 g	3 5.00 g	3 100.00 g

Date: 4/14/2025 Temperature: 20.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.01 g	3 4.97 g	3 99.99 g

Date: 4/22/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 4.99 g	1 99.96 g
2 1.00 g	2 4.99 g	2 99.95 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 4/28/2025 Temperature: 21.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 0.99 g	3 5.02 g	3 99.97 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD3

Monthly Check

Date: 4/1/2025 Temp: 20.6 °C
Analyst: DN Humidity: 65 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.01 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.00 g	3 5.00 g	3 4.99 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 4/7/2025 Temperature: 20.2 °C Weight Set Used: 9072
Analyst: DS Humidity: 39 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 100.00 g
2 1.00 g	2 5.00 g	2 100.00 g
3 1.00 g	3 4.99 g	3 100.00 g

Date: 4/14/2025 Temperature: 20.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.02 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.01 g	3 5.00 g	3 99.97 g

Date: 4/22/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.95 g

Date: 4/28/2025 Temperature: 21.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.02 g	1 99.96 g
2 1.00 g	2 4.99 g	2 99.96 g
3 1.01 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD4

Monthly Check

Date: 4/1/2025 Temp: 20.6 °C
Analyst: DN Humidity: 65 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.99 g	1 5.00 g	1 4.97 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.00 g	2 30.00 g	2 29.99 g
3 1.00 g	3 0.98 g	3 5.00 g	3 5.01 g	3 30.00 g	3 29.99 g

Surrogate Checks

Date: 4/7/2025 Temperature: 20.2 °C Weight Set Used: 9072
Analyst: DS Humidity: 39 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.01 g	1 100.02 g
2 1.01 g	2 5.00 g	2 100.00 g
3 0.99 g	3 4.99 g	3 99.99 g

Date: 4/14/2025 Temperature: 20.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.01 g	1 99.98 g
2 1.00 g	2 5.01 g	2 99.96 g
3 0.99 g	3 4.99 g	3 99.96 g

Date: 4/22/2025 Temperature: 20.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 71 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.00 g	1 99.96 g
2 0.99 g	2 4.99 g	2 99.97 g
3 1.00 g	3 5.01 g	3 99.97 g

Date: 4/28/2025 Temperature: 21.4 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.93 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: M-Toledo

Monthly Check

Date: 4/1/2025 Temp: 20.3 °C
Analyst: DN Humidity: 66 %
Weight Set Used: B323405602

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set B323405602 was used for the weekly verification - DS

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg
2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg
3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg

Surrogate Checks

Date: 4/7/2025 Temperature: 19.7 °C Weight Set Used: B323405602
Analyst: DS Humidity: 43 %

Low Weight Used: 2.00 kg		Mid Weight Used: 2.00 kg		High Weight Used: 2.00 kg	
Balance Readout	1 2.00 kg	Balance Readout	1 2.00 kg	Balance Readout	1 2.00 kg
	2 2.00 kg		2 2.00 kg		2 2.00 kg
	3 2.00 kg		3 2.00 kg		3 2.00 kg

Date: 4/14/2025 Temperature: 19.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 4/22/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 4/28/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: DN Humidity: 66 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: PJ6000

Monthly Check

Date: 4/1/2025 Temp: 20.9 °C
Analyst: DN Humidity: 64 %
Weight Set Used: 9072

Notes: On 04/07/2025 - the surrogate check weight set (155617) was out for its annual calibration, weight set 9072 was used for the weekly verification - DS

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0 g	1	5.00 g	5.0 g	1	30.00 g	30.0 g
2	1.00 g	1.0 g	2	5.00 g	5.0 g	2	30.00 g	30.0 g
3	1.00 g	1.0 g	3	5.00 g	5.0 g	3	30.00 g	30.0 g

Surrogate Checks

Date: 4/7/2025 Temperature: 21.5 °C Weight Set Used: 9072
Analyst: DS Humidity: 40 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 4/14/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 62 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 4/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 4/28/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g