Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Monthly Check

Date: 8/1/2025 Temp: 21.9 °C
Analyst: KB Humidity: 63 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 0.9999 g	1 3.00 g	1 2.9998 g	1 5.00 g	1 5.0004 g
2 1.00 g	2 1.0000 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 5.0000 g
3 1.00 g	3 0.9999 g	3 3.00 g	3 2.9999 g	3 5.00 g	3 5.0002 g

Surrogate Checks

Date: 8/4/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0013 g	1 5.0052 g	1 99.9598 g
2 1.0010 g	2 5.0052 g	2 99.9599 g
3 1.0011 g	3 5.0053 g	3 99.9598 g

Date: 8/11/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0053 g	1 99.9599 g
2 1.0010 g	2 5.0053 g	2 99.9598 g
3 1.0009 g	3 5.0052 g	3 99.9599 g

Date: 8/18/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0011 g	1 5.0053 g	1 99.9600 g
2 1.0010 g	2 5.0053 g	2 99.9599 g
3 1.0010 g	3 5.0053 g	3 99.9599 g

Date: 8/25/2025 Temperature: 21.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0053 g	1 99.9599 g
2 1.0011 g	2 5.0053 g	2 99.9601 g
3 1.0010 g	3 5.0054 g	3 99.9599 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD1

Monthly Check

Date: 8/1/2025 Temp: 22.1 °C
Analyst: KB Humidity: 62 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.01 g	1	5.00 g	5.00 g	1	30.00 g	29.99 g
2	1.00 g	1.00 g	2	5.00 g	4.99 g	2	30.00 g	30.01 g
3	1.00 g	0.99 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 8/4/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 8/11/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 0.99 g	2 5.00 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 8/18/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 8/25/2025 Temperature: 21.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.95 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD2

Monthly Check

Date: 8/1/2025 Temp: 22.1 °C
Analyst: KB Humidity: 62 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.02 g	1	30.00 g	29.99 g
2	1.00 g	0.98 g	2	5.00 g	5.02 g	2	30.00 g	30.02 g
3	1.00 g	1.01 g	3	5.00 g	5.01 g	3	30.00 g	30.02 g

Surrogate Checks

Date: 8/4/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
	Balance Readout		Balance Readout		Balance Readout
1	1.00 g	1	5.01 g	1	99.96 g
2	1.01 g	2	5.01 g	2	99.95 g
3	0.99 g	3	5.00 g	3	99.96 g

Date: 8/11/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
	Balance Readout		Balance Readout		Balance Readout
1	1.00 g	1	5.00 g	1	99.97 g
2	1.00 g	2	5.00 g	2	99.97 g
3	1.00 g	3	5.00 g	3	99.95 g

Date: 8/18/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 74 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
	Balance Readout		Balance Readout		Balance Readout
1	1.00 g	1	5.00 g	1	99.96 g
2	1.00 g	2	5.00 g	2	99.97 g
3	1.00 g	3	4.99 g	3	99.95 g

Date: 8/25/2025 Temperature: 21.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
	Balance Readout		Balance Readout		Balance Readout
1	1.00 g	1	5.01 g	1	99.96 g
2	1.01 g	2	5.00 g	2	99.97 g
3	1.00 g	3	5.02 g	3	99.97 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD3

Monthly Check

Date: 8/1/2025 Temp: 22.1 °C
Analyst: KB Humidity: 62 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	0.98 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	0.98 g	2	5.00 g	4.99 g	2	30.00 g	29.99 g
3	1.00 g	0.99 g	3	5.00 g	5.01 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 8/4/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.02 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 8/11/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.97 g

Date: 8/18/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.97 g
2 1.00 g	2 4.99 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 8/25/2025 Temperature: 21.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.97 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD4

Monthly Check

Date: 8.1.2025 Temp: 22.1 °C
Analyst: KB Humidity: 63 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.02 g	1	5.00 g	5.01 g	1	30.00 g	29.97 g
2	1.00 g	1.03 g	2	5.00 g	5.03 g	2	30.00 g	30.02 g
3	1.00 g	1.03 g	3	5.00 g	5.00 g	3	30.00 g	30.01 g

Surrogate Checks

Date: 8/4/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.98 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.95 g

Date: 8/11/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.01 g	3 5.00 g	3 99.96 g

Date: 8/18/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 4.99 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.94 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 8/25/2025 Temperature: 21.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: M-Toledo

Monthly Check

Date: 8/1/2025 Temp: 22.1 °C
Analyst: KB Humidity: 62 %
Weight Set Used: B323405602

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg
2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg
3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg

Surrogate Checks

Date: 8/4/2025 Temperature: 21.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 8/11/2025 Temperature: 21.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 68 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 8/18/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50

Date: 8/25/2025 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: PJ6000

Monthly Check

Date: 8/1/2025 Temp: 20.3 °C
Analyst: KB Humidity: 68 %
Weight Set Used: 9072

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0 g	1	5.00 g	5.0 g	1	30.00 g	30.0 g
2	1.00 g	1.0 g	2	5.00 g	5.0 g	2	30.00 g	30.0 g
3	1.00 g	1.0 g	3	5.00 g	5.0 g	3	30.00 g	30.0 g

Surrogate Checks

Date: 8/4/2025 Temperature: 19.8 °C Weight Set Used: 155617
Analyst: DN Humidity: 77 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 8/11/2025 Temperature: 18.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 79 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 8/18/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 79 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g

Date: 8/25/2025 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: DN Humidity: 78 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0 g	1 5.0 g	1 100.0 g
2 1.0 g	2 5.0 g	2 100.0 g
3 1.0 g	3 5.0 g	3 100.0 g