Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Monthly Check

Date: 9/2/2025 Temp: 21.2 °C
Analyst: DN Humidity: 74 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.0000 g	1 3.00 g	1 3.0001 g	1 5.00 g	1 5.0001 g
2 1.00 g	2 1.0001 g	2 3.00 g	2 3.0000 g	2 5.00 g	2 5.0001 g
3 1.00 g	3 1.0000 g	3 3.00 g	3 3.0000 g	3 5.00 g	3 5.0001 g

Surrogate Checks

Date: 9/8/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9598 g
2 1.0010 g	2 5.0055 g	2 99.9598 g
3 1.0010 g	3 5.0053 g	3 99.9598 g

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0054 g	1 99.9600 g
2 1.0010 g	2 5.0053 g	2 99.9599 g
3 1.0010 g	3 5.0053 g	3 99.9599 g

Date: 9/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0055 g	1 99.9601 g
2 1.0010 g	2 5.0053 g	2 99.9601 g
3 1.0010 g	3 5.0054 g	3 99.9600 g

Date: 9/29/2025 Temperature: 20.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9601 g
2 1.0010 g	2 5.0051 g	2 99.9598 g
3 1.0010 g	3 5.0053 g	3 99.9598 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD1

Monthly Check

Date: 9/2/2025 Temp: 21.2 °C
Analyst: DN Humidity: 74 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.00 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 9/8/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 9/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 9/29/2025 Temperature: 20.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD2

Monthly Check

Date: 9/2/2025 Temp: 21.2 °C
Analyst: DN Humidity: 74 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.01 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.00 g	2 30.00 g	2 30.01 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.01 g

Surrogate Checks

Date: 9/8/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.01 g	2 4.99 g	2 99.96 g
3 1.01 g	3 5.00 g	3 99.97 g

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 4.99 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 9/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.01 g	1 99.96 g
2 1.00 g	2 4.99 g	2 99.95 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 9/29/2025 Temperature: 20.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD3

Monthly Check

Date: 9/2/2025 Temp: 21.2 °C
Analyst: DN Humidity: 74 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.01 g	2 30.00 g	2 29.99 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 9/8/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.97 g
3 1.01 g	3 5.00 g	3 99.96 g

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 0.99 g	3 5.01 g	3 99.96 g

Date: 9/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.01 g	3 99.96 g

Date: 9/29/2025 Temperature: 20.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 0.99 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.95 g
3 1.01 g	3 5.01 g	3 99.95 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD4

Monthly Check

Date: 9/2/2025 Temp: 21.2 °C
Analyst: DN Humidity: 74 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.01 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.00 g	2 5.00 g	2 5.00 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.00 g

Surrogate Checks

Date: 9/8/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.95 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.97 g

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 4.99 g	1 99.96 g
2 1.01 g	2 5.00 g	2 99.96 g
3 0.99 g	3 5.00 g	3 99.96 g

Date: 9/22/2025 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 4.98 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.94 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 9/29/2025 Temperature: 20.2 °C Weight Set Used: 155617
Analyst: DN Humidity: 65 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.01 g	1 99.97 g
2 1.01 g	2 5.01 g	2 99.96 g
3 0.99 g	3 5.00 g	3 99.97 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: M-Toledo

Monthly Check

Date: 9/2/2025 Temp: 21.4 °C
Analyst: DN Humidity: 71 %
Weight Set Used: B323405602

Notes:

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg
2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg
3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg

Surrogate Checks

Date: 9/8/2025 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: DN Humidity: 66 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 kg

Date: 9/15/2025 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: DN Humidity: 69 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 9/22/2025 Temperature: 20.5 °C Weight Set Used: 155617
Analyst: DN Humidity: 72 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 9/29/2025 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: DN Humidity: 67 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: PJ6000

Monthly Check

Date: 9/2/2025 Temp: 19.5 °C
Analyst: DN Humidity: 81 %
Weight Set Used: 9072

Notes:

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.0 g	1 5.00 g	1 5.0 g	1 30.00 g	1 30.0 g
2 1.00 g	2 1.0 g	2 5.00 g	2 5.0 g	2 30.00 g	2 30.0 g
3 1.00 g	3 1.0 g	3 5.00 g	3 5.0 g	3 30.00 g	3 30.0 g

Surrogate Checks

Date	Analyst	Temperature	Humidity	Weight Set Used
9/8/2025	DN	19.3 °C	75 %	155617
9/15/2025	DN	19.6 °C	79 %	155617
9/22/2025	DN	19.5 °C	77 %	155617
9/29/2025	DN	19.3 °C	72 %	155617

Low Weight Used	Balance Readout	Mid Weight Used	Balance Readout	High Weight Used	Balance Readout
1.00 g	1 1.0 g 2 1.0 g 3 1.0 g	5.00 g	1 5.0 g 2 5.0 g 3 5.0 g	100.00 g	1 100.0 g 2 100.0 g 3 100.0 g