Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: PJ6000

Monthly Check

Date: 5/1/2026 Temp: 19.6 °C
Analyst: KK Humidity: 68 %
Weight Set Used: 9072

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0 g	1	5.00 g	5.0 g	1	30.00 g	30.0 g
2	1.00 g	1.0 g	2	5.00 g	5.0 g	2	30.00 g	30.0 g
3	1.00 g	1.0 g	3	5.00 g	5.0 g	3	30.00 g	30.0 g

Surrogate Checks

Date: 5/4/2026 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK Humidity: 64 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 100.0 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 99.9 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Date: 5/11/2026 Temperature: 19.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 73 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 100.0 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 99.9 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Date: 5/18/2026 Temperature: 19.6 °C Weight Set Used: 155617
Analyst: KK Humidity: 74 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 99.9 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 99.9 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 99.9 g

Date: 5/26/2026 Temperature: 19.9 °C Weight Set Used: 155617
Analyst: KK Humidity: 75 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.0 g	Balance Readout 1: 5.0 g	Balance Readout 1: 99.9 g
Balance Readout 2: 1.0 g	Balance Readout 2: 5.0 g	Balance Readout 2: 100.0 g
Balance Readout 3: 1.0 g	Balance Readout 3: 5.0 g	Balance Readout 3: 100.0 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Page accidentally printed. -KK 6/1/2026

Monthly Check

Date: 4/14/2026 Temp: 20.5 °C
Analyst: KK Humidity: 68 %
Weight Set Used: 9072

Notes: weights were picked up by DS and monthly checks were done the same day. -KK 4/14/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0000 g	1	3.00 g	2.9999 g	1	5.00 g	4.9999 g
2	1.00 g	0.9999 g	2	3.00 g	3.0000 g	2	5.00 g	5.0001 g
3	1.00 g	1.0000 g	3	3.00 g	3.0000 g	3	5.00 g	5.0001 g

Surrogate Checks

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Date: Temperature: °C Weight Set Used:
Analyst: Humidity: %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: g	Balance Readout 1: g	Balance Readout 1: g
Balance Readout 2: g	Balance Readout 2: g	Balance Readout 2: g
Balance Readout 3: g	Balance Readout 3: g	Balance Readout 3: g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: AX224N

Monthly Check

Date: 5/1/2026 Temp: 20.7 °C
Analyst: KK Humidity: 65 %
Weight Set Used: 9072

Balance re-adjusted to center bubble 5/18/2026- KK 5/18/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.0000 g	1	3.00 g	2.9999 g	1	5.00 g	5.0001 g
2	1.00 g	1.0000 g	2	3.00 g	2.9999 g	2	5.00 g	5.0000 g
3	1.00 g	1.0000 g	3	3.00 g	3.0000 g	3	5.00 g	5.0001 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9598 g
2 1.0008 g	2 5.0053 g	2 99.9597 g
3 1.0010 g	3 5.0052 g	3 99.9597 g

Date: 5/11/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0007 g	1 5.0052 g	1 99.9597 g
2 1.0009 g	2 5.0051 g	2 99.9597 g
3 1.0010 g	3 5.0052 g	3 99.9597 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0010 g	1 5.0052 g	1 99.9599 g
2 1.0012 g	2 5.0054 g	2 99.9599 g
3 1.0009 g	3 5.0052 g	3 99.9597 g

Date: 5/26/2026 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.0009 g	1 5.0052 g	1 99.9599 g
2 1.0010 g	2 5.0052 g	2 99.9599 g
3 1.0009 g	3 5.0052 g	3 99.9598 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD1

Monthly Check

Date: 5/1/2026 Temp: 20.7 °C
Analyst: KK Humidity: 65 %
Weight Set Used: 9072

Low		Mid-Level		High	
Weight Used	Balance Readout	Weight Used	Balance Readout	Weight Used	Balance Readout
1 1.00 g	1 1.00 g	1 5.00 g	1 5.00 g	1 30.00 g	1 30.00 g
2 1.00 g	2 1.01 g	2 5.00 g	2 5.00 g	2 30.00 g	2 30.00 g
3 1.00 g	3 1.00 g	3 5.00 g	3 5.00 g	3 30.00 g	3 30.02 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.01 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.98 g

Date: 5/11/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.00 g	2 99.96 g
3 1.00 g	3 5.00 g	3 99.96 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.97 g
2 1.00 g	2 5.00 g	2 99.95 g
3 1.00 g	3 5.01 g	3 99.97 g

Date: 5/26/2026 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.00 g	1 99.94 g
2 1.00 g	2 5.00 g	2 99.95 g
3 1.00 g	3 5.00 g	3 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD2

Monthly Check

Date: 5/1/2026 Temp: 20.7 °C
Analyst: KK Humidity: 65 %
Weight Set Used: 9072

Bubble was off-center, balance re-adjusted. -KK 5/11/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	5.00 g	1	30.00 g	30.00 g
2	1.00 g	0.99 g	2	5.00 g	5.00 g	2	30.00 g	30.01 g
3	1.00 g	1.01 g	3	5.00 g	4.99 g	3	30.00 g	30.01 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 0.99 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.98 g
Balance Readout 2: 1.02 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Date: 5/11/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.01 g	Balance Readout 1: 99.97 g
Balance Readout 2: 0.99 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.95 g
Balance Readout 3: 0.99 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.94 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.01 g	Balance Readout 1: 4.99 g	Balance Readout 1: 99.97 g
Balance Readout 2: 1.00 g	Balance Readout 2: 5.01 g	Balance Readout 2: 99.97 g
Balance Readout 3: 1.00 g	Balance Readout 3: 5.00 g	Balance Readout 3: 99.93 g

Date: 5/26/2026 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout 1: 1.00 g	Balance Readout 1: 5.00 g	Balance Readout 1: 99.94 g
Balance Readout 2: 1.01 g	Balance Readout 2: 5.00 g	Balance Readout 2: 99.96 g
Balance Readout 3: 1.02 g	Balance Readout 3: 5.01 g	Balance Readout 3: 99.96 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD3

Monthly Check

Date: 5/1/2026 Temp: 20.7 °C
Analyst: KK Humidity: 65 %
Weight Set Used: 9072

Banance re-adjusted to center bubble -KK
5/18/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.02 g	1	5.00 g	5.00 g	1	30.00 g	29.99 g
2	1.00 g	1.00 g	2	5.00 g	5.00 g	2	30.00 g	30.02 g
3	1.00 g	1.00 g	3	5.00 g	5.02 g	3	30.00 g	29.99 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.00 g	2 5.01 g	2 99.96 g
3 0.99 g	3 5.00 g	3 99.97 g

Date: 5/11/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.96 g
2 1.01 g	2 5.01 g	2 99.97 g
3 1.01 g	3 5.00 g	3 99.95 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.00 g	1 5.01 g	1 99.95 g
2 1.00 g	2 5.00 g	2 99.94 g
3 0.99 g	3 5.03 g	3 99.95 g

Date: 5/26/2026 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 1.00 g	Mid Weight Used: 5.00 g	High Weight Used: 100.00 g
Balance Readout	Balance Readout	Balance Readout
1 1.01 g	1 5.00 g	1 99.96 g
2 1.00 g	2 5.02 g	2 99.95 g
3 0.99 g	3 4.98 g	3 99.95 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: SD4

Monthly Check

Date: 5/1/2026 Temp: 20.7 °C
Analyst: KK Humidity: 65 %
Weight Set Used: 9072

Bubble was off-center, balance re-adjusted. -KK 5/11/2026

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	1.00 g	1.00 g	1	5.00 g	4.99 g	1	30.00 g	30.01 g
2	1.00 g	0.98 g	2	5.00 g	5.00 g	2	30.00 g	30.01 g
3	1.00 g	1.01 g	3	5.00 g	5.00 g	3	30.00 g	30.00 g

Surrogate Checks

Date: 5/4/2026 Temperature: 20.1 °C Weight Set Used: 155617
Analyst: KK Humidity: 59 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.01 g	Balance Readout	1 5.02 g	Balance Readout	1 99.95 g
	2 1.00 g		2 5.01 g		2 99.95 g
	3 1.00 g		3 5.00 g		3 99.96 g

Date: 5/11/2026 Temperature: 20.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 69 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.01 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.01 g		2 99.97 g
	3 1.00 g		3 5.00 g		3 99.96 g

Date: 5/18/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 68 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.00 g	Balance Readout	1 5.00 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.00 g		2 99.96 g
	3 1.00 g		3 5.00 g		3 99.96 g

Date: 5/26/2026 Temperature: 21.3 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 1.00 g		Mid Weight Used: 5.00 g		High Weight Used: 100.00 g	
Balance Readout	1 1.01 g	Balance Readout	1 5.00 g	Balance Readout	1 99.96 g
	2 1.00 g		2 5.01 g		2 99.96 g
	3 1.00 g		3 5.01 g		3 99.95 g

Brazoria County Crime Laboratory

Balance Monitoring and Maintenance

Balance Identification: M-Toledo

Monthly Check

Date: 5/1/2026 Temp: 20.3 °C
Analyst: KK Humidity: 68 %
Weight Set Used: B323405602

Low			Mid-Level			High		
	Weight Used	Balance Readout		Weight Used	Balance Readout		Weight Used	Balance Readout
1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg	1	2.00 kg	2.00 kg
2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg	2	2.00 kg	2.00 kg
3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg	3	2.00 kg	2.00 kg

Surrogate Checks

Date: 5/4/2026 Temperature: 19.8 °C Weight Set Used: 155617
Analyst: KK Humidity: 66 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 5/11/2026 Temperature: 20.7 °C Weight Set Used: 155617
Analyst: KK Humidity: 68 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 5/18/2026 Temperature: 20.9 °C Weight Set Used: 155617
Analyst: KK Humidity: 67 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg

Date: 5/26/2026 Temperature: 21.6 °C Weight Set Used: 155617
Analyst: KK Humidity: 70 %

Low Weight Used: 0.50 kg		Mid Weight Used: 0.50 kg		High Weight Used: 0.50 kg	
Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg	Balance Readout	1 0.50 kg
	2 0.50 kg		2 0.50 kg		2 0.50 kg
	3 0.50 kg		3 0.50 kg		3 0.50 kg