BRAZORIA COUNTY SHERIFF'S OFFICE
CRIME LABORATORY
3602 County Road 45, Angleton, TX 77515
(979) 864-2250 • CrimeLab@brazoriacountytx.gov

SEIZED DRUG MEASUREMENT UNCERTAINTY

Balance Case Number: MU SD2
Report Number: 0003
Report Date: 5/9/2025

Balance: Ohaus AX2202 (SD2)
SN: C144891582
Effective Dates: 5/11/2025 to 11/11/2026

RESULTS:

	u (k=2)	u (k=3)
Reported Uncertainty:	0.03 g	0.04 g

File Attachment(s):

- UoM 2025_Ohaus AX2202 SD2

Laboratory Policy:

All weight measurements shall be reported at the 95% coverage probability, k=2, rounded to the second decimal place to correlate to the balance resolution.

Analyst: _____
Aleia Winters