July 18, 2024

Analyst: Melissa Henry
Area of Proficiency Tested: Toxicology – Blood Alcohol
Date of Test Completion: March 21, 2024
Proficiency Test Service Provider: Collaborative Testing Service
Source of Test: External
Test Number: CTS-24-5641
Lab Number: BCCL-0221

This letter discusses your performance on the proficiency examination identified above that was completed as part of your competency testing. Your findings were satisfactory and in line with the collaborative group findings in the summary report. The results are as follows:

ITEM 1: Your calculated mean was 0.084 g/100mL ± 0.012 g/100mL.
The collaborative group's grand mean is 0.0818 g/100mL with a standard deviation of 0.0023 g/100mL according to the report summary.

ITEM 2: Your calculated mean was 0.260 g/100mL ± 0.037 g/100mL.
The collaborative group's grand mean is 0.2624 g/100mL with a standard deviation of 0.0063 g/100mL according to the report summary.

ITEM 3: Your calculated mean was 0.315 g/100mL ± 0.045 g/100mL.
The collaborative group's grand mean is 0.3027 g/100mL with a standard deviation of 0.0079 g/100mL according to the report summary.

ITEM 4: Your calculated mean was 0.199 g/100mL ± 0.029 g/100mL.
The collaborative group's grand mean is 0.1942 g/100mL with a standard deviation of 0.0051 g/100mL according to the report summary.

Aleia Winters, Laboratory Quality Manager

Acknowledged: Melissa Henry Date: 7/18/2024