December 23, 2025

Analyst: Kayla Baylor
Area of Proficiency Tested: Toxicology – Blood Alcohol
Date of Test Completion: October 13, 2025
Proficiency Test Service Provider: Collaborative Testing Service
Source of Test: External
Test Number: CTS-25-5642
Lab Number: BCCL-25-1651

This letter discusses your performance on the proficiency examination identified above that was completed as part of your competency testing. Your findings were satisfactory and in line with the collaborative group findings in the summary report. The results are as follows:

ITEM 1: Your calculated mean was 0.265 g/100mL ± 0.043 g/100mL.
The collaborative group's grand mean is 0.2892 g/100mL with a standard deviation of 0.0082 g/100mL according to the report summary.

ITEM 2: Your calculated mean was 0.156 g/100mL ± 0.025 g/100mL.
The collaborative group's grand mean is 0.1693 g/100mL with a standard deviation of 0.0052 g/100mL according to the report summary.

ITEM 3: Your calculated mean was 0.284 g/100mL ± 0.046 g/100mL.
The collaborative group's grand mean is 0.3122 g/100mL with a standard deviation of 0.0097 g/100mL according to the report summary.

ITEM 4: Your calculated mean was 0.063 g/100mL ± 0.010 g/100mL.
The collaborative group's grand mean is 0.0688 g/100mL with a standard deviation of 0.0032 g/100mL according to the report summary.

Aleia Winters Digitally signed by Aleia Winters Date: 2026.01.07 06:56:20 -06'00'

Aleia Winters, Laboratory Quality Manager

Acknowledged: Kayla M. Baylor Digitally signed by Kayla M. Baylor Date: 2026.01.05 12:55:25 -06'00' Date: 1/5/2025